

EC – DECLARATION OF CONFORMITY

The product named hereinafter was developed, designed, and manufactured in compliance with the relevant, fundamental safety and health requirements of the listed EC directives and norms. In the event of modifications that were not authorised by us or if the product is used in a manner that is not in line with the intended purpose, this declaration will be rendered void.

| | |
|----------------------|---|
| <i>Product name:</i> | Laboratory centrifuge |
| <i>Product type:</i> | Sigma 6-16S, 6-16HS |
| <i>Order number:</i> | 10390, 10391, 10392, 10393 |
| <i>Directives:</i> | 2006/42/EC Machinery Directive 2014/35/EU Low Voltage Directive 2014/30/EU EMC Directive 2011/65/EU RoHS Directive |
| <i>Normes:</i> | EN 61010-2-020:2006 EN 61000-3-2:2014 EN 61000-3-3:2013 EN 61326-1:2013 |

| | |
|--|--|
| <p>Sigma Laborzentrifugen GmbH An der Unteren Söse 50 37520 Osterode Germany</p> <p>Osterode, 13/09/2017</p> <p><i>Michael Sander</i> General Manager</p> | <p>Authorised representative for CE matters: Eckhard Tödteberg</p> |
|--|--|