

## PRODUCT DATA SHEET

Standing: 2021-12-06

### LAUDA Varioshake VS 8 O

Shaker 230 V; 50/60 Hz

Part Number: L003057

#### Features

- Shaker with orbital motion, heavy duty, low noise, in a compact design
- Shaking frequency display and setting digitally via LC display, in steps of 1.0 rpm, gentle start-up
- The setpoint and actual values as well as the shaking frequency are displayed on two LC displays
- User-friendly microprocessor technology
- Membrane keyboard with distinctive symbols.
- Timer adjustable from 1 min. up to 99:59 hours or continuous operation
- Optional RS 232 interface on request
- Shaking table made of anodized aluminum with four plastic pins
- Drive by an overload-protected AC motor
- Compact, low-wear mechanism with mass balancing for a secure stand
- Constant speed, independent of load
- Excellent for continuous use
- Outer housing made of electrolytically galvanized sheet steel, powder-coated



Reserve technical changes

#### Technical Features

Ambient temperature range	10 ... 50 °C
Shaking tray (WxD)	330 x 330 mm
Power consumption max.	0.07 kW
Shaking amplitude	10 mm
Shaking frequency	20 ... 500 1/min
Load max.	8 kg
Overall dimensions (WxDxH)	350 x 355 x 160 mm
Power plug	Power cord with angled plug (CEE7/7)
Power supply	230 V; 50/60 Hz

LAUDA DR. R. WOBSEY GMBH & CO. KG  
Laudaplatz 1 • 97922 Lauda-Königshofen  
Postfach 1251 • 97912 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0 • F + 49 (0) 9343 503-222  
info@lauda.de • www.lauda.de  
WEEE-Reg-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen  
Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:  
LAUDA DR. R. WOBSEY Verwaltungs-GmbH  
Sitz Lauda-Königshofen  
Registergericht Mannheim • HRB 560226

Geschäftsführer:  
Dr. Gunther Wobser (Vors.), Dr. Mario Englert,  
Dr. Marc Stricker  
Beirat: Dr. Gerhard Wobser